

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE041720\
 Data File : FE031774.D
 Signal(s) : FID1B.CH
 Acq On : 17 Apr 2020 17:30
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Apr 18 00:51:26 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE040720.M
 Quant Title :
 QLast Update : Tue Apr 07 15:06:47 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.388	17744337	48.722 ug/ml
Target Compounds			
2) N-DECANE	4.156	20012445	51.788 ug/ml
3) N-DODECANE	6.158	20246615	54.046 ug/ml
4) N-TETRADECANE	7.948	21254909	53.625 ug/ml
5) N-HEXADECANE	9.546	20907230	51.268 ug/ml
6) N-OCTADECANE	10.985	21507178	50.650 ug/ml
7) N-EICOSANE	12.293	20914997	49.693 ug/ml
8) N-DOCOSANE	13.490	20821880	49.109 ug/ml
10) N-TETRACOSANE	14.594	20317770	48.538 ug/ml
11) N-HEXACOSANE	15.614	19902497	48.025 ug/ml
12) N-OCTACOSANE	16.563	19708138	47.904 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE041720\
Data File : FE031774.D
Signal(s) : FID1B.CH
Acq On : 17 Apr 2020 17:30
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Apr 18 00:51:26 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE040720.M
Quant Title :
QLast Update : Tue Apr 07 15:06:47 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

