

rteres

Instrument :

FID_E

LabSampleId :

20 PPM ALIPHATIC HC STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 04/18/2023

Supervised By :Ankita Jodhani 04/18/2023

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE04172
Data File : FE041397.D
Signal(s) : FID1B.ch
Acq On : 17 Apr 2023 23:42
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 032523.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.237	3.207	3.350	BB	323055	3285217	76.88%	4.504%
2	4.484	4.444	4.592	BB	319581	3266625	76.45%	4.478%
3	6.187	6.160	6.387	BB	305163	3563077	83.38%	4.885%
4	6.649	6.607	6.746	BB	336351	3352312	78.45%	4.596%
5	7.284	7.244	7.462	BB	275862	3311844	77.50%	4.540%
6	8.476	8.432	8.577	BB	313852	3373198	78.94%	4.624%
7	10.083	10.037	10.176	BB	316094	3433279	80.35%	4.707%
8	11.526	11.488	11.595	BB	299103	3431837	80.31%	4.705%
9	11.840	11.793	11.925	BB	321404	3564835	83.42%	4.887%
10	12.836	12.748	12.902	BB	280759	3274889	76.64%	4.489%
11	13.281	13.230	13.355	BB	232935	2748446	64.32%	3.768%
12	13.448	13.373	13.517	BV	269422	3239649	75.81%	4.441%
13	14.036	14.004	14.107	VB	271402	3189560	74.64%	4.372%
14	15.141	15.067	15.212	BB	260613	3133971	73.34%	4.296%
15	16.164	15.700	16.206	PB	242734	2976134	69.65%	4.080%
16	17.115	16.694	17.173	BV	229279	3103011	72.62%	4.254%
17	18.005	17.598	18.045	PV	225446	2879706	67.39%	3.948%
18	18.839	18.751	18.877	BV	232857	3232980	75.66%	4.432%
19	19.626	19.561	19.662	BV	223879	3208961	75.10%	4.399%
20	20.369	20.180	20.401	PV	215558	3871488	90.60%	5.307%
21	21.095	21.054	21.137	VV	195313	3232261	75.64%	4.431%
22	21.979	21.895	22.060	VV	163138	4273130	100.00%	5.858%
Sum of corrected areas:							72946412	

Aliphatic EPH 032523.M Tue Apr 18 01:32:50 2023