

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE041723AL\
 Data File : FE041395.D
 Signal(s) : FID1B.ch
 Acq On : 17 Apr 2023 22:11
 Operator : YP\AJ
 Sample : 02370-02
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 HR-04-041423-EPH-2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/18/2023
 Supervised By :Ankita Jodhani 04/18/2023

Integration File: autoint1.e
 Quant Time: Apr 18 01:04:00 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 032523.M
 Quant Title : GC Extractables
 QLast Update : Sat Mar 25 06:32:15 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.844	6107857	32.499 ug/mlm
Spiked Amount	50.000	Recovery	= 65.00%
12) S 1-chlorooctadecane (S...	13.334	213745751	1485.586 ug/ml
Spiked Amount	50.000	Recovery	= 2971.17%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE041723AL\
Data File : FE041395.D
Signal(s) : FID1B.ch
Acq On : 17 Apr 2023 22:11
Operator : YP\AJ
Sample : 02370-02
Misc :
ALS Vial : 26 Sample Multiplier: 1


Instrument :
FID_E
ClientSampleId :
HR-04-041423-EPH-2


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/18/2023
Supervised By :Ankita Jodhani 04/18/2023

Integration File: autoint1.e
Quant Time: Apr 18 01:04:00 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 032523.
Quant Title : GC Extractables
QLast Update : Sat Mar 25 06:32:15 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.844 min
Delta R.T.: -0.021 min
Response: 6107857
Conc: 32.50 ug/ml

Instrument :
FID_E
ClientSampleId :
HR-04-041423-EPH-2

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/18/2023
Supervised By :Ankita Jodhani 04/18/2023



#12 1-chlorooctadecane (SURR)

R.T.: 13.334 min
Delta R.T.: 0.032 min
Response: 213745751
Conc: 1485.59 ug/ml