

rteres

Instrument :
FID_E
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE04192
Data File : FE041427.D
Signal(s) : FID1B.ch
Acq On : 19 Apr 2023 14:07
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 04/20/2023
Supervised By :Ankita Jodhani 04/20/2023

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 032523.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.232	3.205	3.339	BB	313205	3185098	89.94%	4.508%
2	4.478	4.441	4.588	BB	314488	3181251	89.83%	4.502%
3	6.182	6.158	6.388	BB	286636	3491643	98.59%	4.942%
4	6.644	6.585	6.752	BB	320898	3268928	92.30%	4.626%
5	7.278	7.249	7.472	PB	257333	3263258	92.14%	4.618%
6	8.470	8.423	8.571	BB	306618	3298443	93.14%	4.668%
7	10.077	10.034	10.162	BB	308550	3394473	95.85%	4.804%
8	11.519	11.482	11.599	PB	306231	3424246	96.69%	4.846%
9	11.833	11.787	11.893	BV	324811	3541475	100.00%	5.012%
10	12.828	12.736	12.907	BB	279433	3282648	92.69%	4.646%
11	13.273	13.205	13.365	BV	234409	2792990	78.87%	3.953%
12	13.440	13.406	13.482	PV	276534	3232374	91.27%	4.575%
13	14.027	13.993	14.101	BB	265261	3172661	89.59%	4.490%
14	15.131	15.041	15.225	BB	245529	3211648	90.69%	4.545%
15	16.154	16.116	16.216	BV	231579	2992080	84.49%	4.235%
16	17.105	17.040	17.169	BB	226521	2981821	84.20%	4.220%
17	17.994	17.799	18.036	BB	224496	3058702	86.37%	4.329%
18	18.828	18.751	18.888	BV	227468	3165828	89.39%	4.480%
19	19.614	19.561	19.654	BV	222080	3156537	89.13%	4.467%
20	20.356	20.021	20.396	PV	211053	2991868	84.48%	4.234%
21	21.082	21.031	21.143	BV	192876	3178904	89.76%	4.499%
22	21.961	21.751	22.095	BB	143695	3391498	95.77%	4.800%
Sum of corrected areas:							70658377	

Aliphatic EPH 032523.M Thu Apr 20 03:20:26 2023