

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_E\Data\FE041923AL\  
 Data File : FE041443.D  
 Signal(s) : FID1B.ch  
 Acq On : 19 Apr 2023 22:10  
 Operator : YP\AJ  
 Sample : 02390-10 10X  
 Misc :  
 ALS Vial : 31 Sample Multiplier: 1

Instrument :  
 FID\_E  
 ClientSampleId :  
 SPC-3WC-23-04-17

Integration File: autoint1.e  
 Quant Time: Apr 20 02:00:34 2023  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_E\methods\Aliphatic EPH 032523.M  
 Quant Title : GC Extractables  
 QLast Update : Sat Mar 25 06:32:15 2023  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 ul  
 Signal Phase : Rxi-1ms  
 Signal Info : 20M x 0.18mm x 0.18um

| Compound                       | R.T.   | Response   | Conc Units  |
|--------------------------------|--------|------------|-------------|
| -----                          |        |            |             |
| System Monitoring Compounds    |        |            |             |
| 9) S ortho-Terphenyl (SURR)    | 11.830 | 427743     | 2.276 ug/ml |
| Spiked Amount 50.000           |        | Recovery = | 4.55%       |
| 12) S 1-chlorooctadecane (S... | 13.270 | 381872     | 2.654 ug/ml |
| Spiked Amount 50.000           |        | Recovery = | 5.31%       |

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_E\Data\FE041923AL\  
Data File : FE041443.D  
Signal(s) : FID1B.ch  
Acq On : 19 Apr 2023 22:10  
Operator : YP\AJ  
Sample : 02390-10 10X  
Misc :  
ALS Vial : 31 Sample Multiplier: 1

Instrument :  
FID\_E  
ClientSampleId :  
SPC-3WC-23-04-17

Integration File: autoint1.e  
Quant Time: Apr 20 02:00:34 2023  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_E\methods\Aliphatic EPH 032523.M  
Quant Title : GC Extractables  
QLast Update : Sat Mar 25 06:32:15 2023  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.830 min  
Delta R.T.: -0.035 min  
Response: 427743  
Conc: 2.28 ug/ml

Instrument :  
FID\_E  
ClientSampleId :  
SPC-3WC-23-04-17



#12 1-chlorooctadecane (SURR)

R.T.: 13.270 min  
Delta R.T.: -0.032 min  
Response: 381872  
Conc: 2.65 ug/ml