

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE042120\
 Data File : FE031777.D
 Signal(s) : FID1B.CH
 Acq On : 21 Apr 2020 11:54
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Apr 22 00:30:36 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE040720.M
 Quant Title :
 QLast Update : Tue Apr 07 15:06:47 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.398	19590728	53.791 ug/ml
Target Compounds			
2) N-DECANE	4.159	21636502	55.991 ug/ml
3) N-DODECANE	6.162	21946882	58.585 ug/ml
4) N-TETRADECANE	7.953	23089975	58.254 ug/ml
5) N-HEXADECANE	9.552	22746416	55.778 ug/ml
6) N-OCTADECANE	10.993	23476187	55.287 ug/ml
7) N-EICOSANE	12.303	22907079	54.426 ug/ml
8) N-DOCOSANE	13.501	22858010	53.911 ug/ml
10) N-TETRACOSANE	14.604	22380681	53.466 ug/ml
11) N-HEXACOSANE	15.625	22000369	53.087 ug/ml
12) N-OCTACOSANE	16.576	21756220	52.882 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE042120\
Data File : FE031777.D
Signal(s) : FID1B.CH
Acq On : 21 Apr 2020 11:54
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Apr 22 00:30:36 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE040720.M
Quant Title :
QLast Update : Tue Apr 07 15:06:47 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

