

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE042120\
 Data File : FE031780.D
 Signal(s) : FID1B.CH
 Acq On : 21 Apr 2020 17:19
 Operator : DD\AJ
 Sample : PB128382BS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Apr 22 00:31:18 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE040720.M
 Quant Title :
 QLast Update : Tue Apr 07 15:06:47 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.386	7517786	20.642 ug/ml
Target Compounds			
2) N-DECANE	4.151	7922986	20.503 ug/ml
3) N-DODECANE	6.152	7874997	21.022 ug/ml
4) N-TETRADECANE	7.942	8614952	21.735 ug/ml
5) N-HEXADECANE	9.540	8499668	20.843 ug/ml
6) N-OCTADECANE	10.981	8764593	20.641 ug/ml
7) N-EICOSANE	12.290	8647734	20.547 ug/ml
8) N-DOCOSANE	13.488	8632568	20.360 ug/ml
10) N-TETRACOSANE	14.592	8446799	20.179 ug/ml
11) N-HEXACOSANE	15.612	8289770	20.003 ug/ml
12) N-OCTACOSANE	16.562	8158078	19.830 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE042120\
Data File : FE031780.D
Signal(s) : FID1B.CH
Acq On : 21 Apr 2020 17:19
Operator : DD\AJ
Sample : PB128382BS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Apr 22 00:31:18 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE040720.M
Quant Title :
QLast Update : Tue Apr 07 15:06:47 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

