

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE042120\
 Data File : FE031812.D
 Signal(s) : FID1B.CH
 Acq On : 22 Apr 2020 15:52
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Apr 22 16:39:40 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE040720.M
 Quant Title :
 QLast Update : Tue Apr 07 15:06:47 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.389	16376313	44.965 ug/ml
Target Compounds			
2) N-DECANE	4.153	17848347	46.188 ug/ml
3) N-DODECANE	6.155	18697692	49.912 ug/ml
4) N-TETRADECANE	7.945	19107535	48.207 ug/ml
5) N-HEXADECANE	9.543	18822498	46.156 ug/ml
6) N-OCTADECANE	10.984	19525101	45.982 ug/ml
7) N-EICOSANE	12.292	19068973	45.307 ug/ml
8) N-DOCOSANE	13.491	19167570	45.207 ug/ml
10) N-TETRACOSANE	14.594	18825891	44.974 ug/ml
11) N-HEXACOSANE	15.615	18381490	44.355 ug/ml
12) N-OCTACOSANE	16.565	18061042	43.901 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE042120\
Data File : FE031812.D
Signal(s) : FID1B.CH
Acq On : 22 Apr 2020 15:52
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 30 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Apr 22 16:39:40 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE040720.M
Quant Title :
QLast Update : Tue Apr 07 15:06:47 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

