

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_E\Data\FE042324AL\  
 Data File : FE047738.D  
 Signal(s) : FID1B.ch  
 Acq On : 23 Apr 2024 15:19  
 Operator : YP\AJ  
 Sample : P2227-02  
 Misc :  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 FID\_E  
 ClientSampleId :  
 HD-02-04222024

Integration File: autoint1.e  
 Quant Time: Apr 23 15:42:35 2024  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_E\methods\Aliphatic EPH 040224.M  
 Quant Title : GC Extractables  
 QLast Update : Tue Apr 02 05:28:28 2024  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 ul  
 Signal Phase : Rxi-1ms  
 Signal Info : 20M x 0.18mm x 0.18um

| Compound                       | R.T.   | Response   | Conc Units   |
|--------------------------------|--------|------------|--------------|
| -----                          |        |            |              |
| System Monitoring Compounds    |        |            |              |
| 9) S ortho-Terphenyl (SURR)    | 11.975 | 4782758    | 27.575 ug/ml |
| Spiked Amount 50.000           |        | Recovery = | 55.15%       |
| 12) S 1-chlorooctadecane (S... | 13.410 | 5495890    | 40.863 ug/ml |
| Spiked Amount 50.000           |        | Recovery = | 81.73%       |

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_E\Data\FE042324AL\  
Data File : FE047738.D  
Signal(s) : FID1B.ch  
Acq On : 23 Apr 2024 15:19  
Operator : YP\AJ  
Sample : P2227-02  
Misc :  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
FID\_E  
ClientSampleId :  
HD-02-04222024

Integration File: autoint1.e  
Quant Time: Apr 23 15:42:35 2024  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_E\methods\Aliphatic EPH 040224.M  
Quant Title : GC Extractables  
QLast Update : Tue Apr 02 05:28:28 2024  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.975 min  
Delta R.T.: -0.007 min  
Response: 4782758  
Conc: 27.57 ug/ml

Instrument :  
FID\_E  
ClientSampleId :  
HD-02-04222024



#12 1-chlorooctadecane (SURR)

R.T.: 13.410 min  
Delta R.T.: -0.005 min  
Response: 5495890  
Conc: 40.86 ug/ml