

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE042721\
 Data File : FE033658.D
 Signal(s) : FID1B.CH
 Acq On : 27 Apr 2021 19:02
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Ankita
 4/28/2021 3:46:53 PM

Integration File: autoint1.e
 Quant Time: Apr 28 07:38:38 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE042221.M
 Quant Title :
 QLast Update : Fri Apr 23 11:28:58 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.751	19330732	48.190 ug/ml
Target Compounds			
1) N-OCTANE	2.422	20716636	49.845 ug/ml
2) N-DECANE	4.484	21954904	50.950 ug/ml
3) N-DODECANE	6.500	22369219	51.304 ug/ml
4) N-TETRADECANE	8.297	22797532	50.704 ug/ml
5) N-HEXADECANE	9.899	22606135	50.058 ug/ml
6) N-OCTADECANE	11.342	22962156	49.552 ug/ml
7) N-EICOSANE	12.653	22606576	49.056 ug/ml
8) N-DOCOSANE	13.853	22355784	48.565 ug/ml
10) N-TETRACOSANE	14.958	21968420	48.007 ug/ml
11) N-HEXACOSANE	15.980	21291001	47.434 ug/ml
12) N-OCTACOSANE	16.931	20703797	47.193 ug/ml
13) N-TRIACONTANE	17.820	20012959	47.269 ug/ml
14) N-DOTRIACONTANE	18.653	19150876	47.374 ug/ml
15) N-TETRATRIACONTANE	19.438	18537557	49.258 ug/mlm
16) N-HEXATRIACONTANE	20.178	17885246	51.069 ug/mlm
17) N-OCTATRIACONTANE	20.930	17229774	53.778 ug/ml
18) N-TETRACONTANE	21.873	17560235	63.260 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE042721\
Data File : FE033658.D
Signal(s) : FID1B.CH
Acq On : 27 Apr 2021 19:02
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_E
ClientSampled :
50 PPM TRPH STD

Manual Integrations
APPROVED

Ankita
4/28/2021 3:46:53 PM

Integration File: autoint1.e
Quant Time: Apr 28 07:38:38 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE042221.M
Quant Title :
QLast Update : Fri Apr 23 11:28:58 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

