

rteres

Instrument :
FID_E
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 05/04/2023
Supervised By :Ankita Jodhani 05/04/2023

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE05032
Data File : FE041720.D
Signal(s) : FID1B.ch
Acq On : 04 May 2023 01:57
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 042123.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.354	3.309	3.408	BB	311242	3027582	80.25%	4.345%
2	4.608	4.566	4.656	BB	318298	3143294	83.32%	4.511%
3	6.316	6.218	6.353	BV	381431	3471449	92.02%	4.982%
4	6.778	6.736	6.820	BB	325952	3188972	84.53%	4.576%
5	7.413	7.357	7.473	BB	356573	3390315	89.87%	4.865%
6	8.607	8.561	8.656	BB	310794	3197809	84.77%	4.589%
7	10.216	10.169	10.266	BB	306477	3326096	88.17%	4.773%
8	11.660	11.611	11.711	BB	300436	3368607	89.29%	4.834%
9	11.981	11.935	12.031	BB	348275	3772514	100.00%	5.414%
10	12.971	12.891	13.011	BB	272950	3229412	85.60%	4.634%
11	13.417	13.371	13.466	BB	236549	2781749	73.74%	3.992%
12	13.584	13.548	13.631	BB	264970	3159777	83.76%	4.534%
13	14.170	14.136	14.221	VB	256089	3133830	83.07%	4.497%
14	15.274	15.195	15.336	BB	254164	3101192	82.20%	4.450%
15	16.297	16.073	16.341	BV	230864	2985507	79.14%	4.284%
16	17.248	17.002	17.288	BB	224446	2917718	77.34%	4.187%
17	18.135	18.056	18.172	PV	222362	3037082	80.51%	4.358%
18	18.967	18.913	19.030	M	217244	2985852	79.15%	4.285%
19	19.752	19.696	19.787	VV	197569	2846184	75.45%	4.084%
20	20.492	20.446	20.536	VV	195548	3237841	85.83%	4.646%
21	21.248	21.190	21.301	PV	169404	3127380	82.90%	4.488%
22	22.193	22.129	22.247	PV	126619	3256223	86.31%	4.673%
Sum of corrected areas:						69686384		

Aliphatic EPH 042123.M Thu May 04 03:09:17 2023