

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE050521\
 Data File : FE033725.D
 Signal(s) : FID1B.CH
 Acq On : 05 May 2021 23:17
 Operator : DD\AJ
 Sample : FE050521ICV
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 FE050521ICV

Integration File: autoint1.e
 Quant Time: May 06 02:21:02 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE050521.M
 Quant Title :
 QLast Update : Thu May 06 02:18:38 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.734	19430081	48.073 ug/ml
Target Compounds			
1) N-OCTANE	2.419	20778530	50.094 ug/ml
2) N-DECANE	4.471	21422519	49.934 ug/ml
3) N-DODECANE	6.485	21953886	50.171 ug/ml
4) N-TETRADECANE	8.281	22401355	49.730 ug/ml
5) N-HEXADECANE	9.884	22378786	49.658 ug/ml
6) N-OCTADECANE	11.326	22615028	49.095 ug/ml
7) N-EICOSANE	12.637	23038476	50.187 ug/ml
8) N-DOCOSANE	13.836	22992593	49.865 ug/ml
10) N-TETRACOSANE	14.943	22736806	49.243 ug/ml
11) N-HEXACOSANE	15.964	22535324	49.114 ug/ml
12) N-OCTACOSANE	16.914	22462946	48.976 ug/ml
13) N-TRIACONTANE	17.804	22307116	48.933 ug/ml
14) N-DOTRIACONTANE	18.637	22104310	49.658 ug/ml
15) N-TETRATRIACONTANE	19.422	21003017	48.860 ug/ml
16) N-HEXATRIACONTANE	20.163	20135851	50.224 ug/ml
17) N-OCTATRIACONTANE	20.909	19213338	52.134 ug/ml
18) N-TETRACONTANE	21.844	18119827	50.110 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE050521\
Data File : FE033725.D
Signal(s) : FID1B.CH
Acq On : 05 May 2021 23:17
Operator : DD\AJ
Sample : FE050521ICV
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
FE050521ICV

Integration File: autoint1.e
Quant Time: May 06 02:21:02 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE050521.M
Quant Title :
QLast Update : Thu May 06 02:18:38 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

