

Method Path : P:\HPCHEM1\FID_E\methods\
 Method File : FE050818.M
 Title :
 Last Update : Wed May 09 08:38:37 2018
 Response Via : Initial Calibration

Calibration Files

50 =FE026985.D 100 =FE026984.D 20 =FE026986.D
 10 =FE026987.D 5 =FE026988.D

	Compound	50	100	20	10	5	Avg		%RSD
1)	N-OCTANE	1.458	1.402	1.467	1.594	1.650	1.514	E5	6.83
2)	N-DECANE	1.582	1.546	1.570	1.694	1.753	1.629	E5	5.51
3)	N-DODECANE	1.740	1.729	1.676	1.765	1.782	1.738	E5	2.34
4)	N-TETRADECANE	1.900	1.892	1.814	1.886	1.958	1.890	E5	2.73
5)	N-HEXADECANE	1.979	1.966	1.898	1.945	2.041	1.966	E5	2.65
6)	N-OCTADECANE	1.985	1.965	1.931	1.996	2.164	2.008	E5	4.50
7)	N-EICOSANE	1.892	1.869	1.840	1.906	2.014	1.904	E5	3.47
8)	N-DOCOSANE	1.775	1.730	1.725	1.823	1.865	1.784	E5	3.39
9) S	TETRACOSANE-d50 (S...	1.462	1.399	1.424	1.536	1.542	1.472	E5	4.39
10)	N-TETRACOSANE	1.667	1.588	1.629	1.768	1.770	1.684	E5	4.87
11)	N-HEXACOSANE	1.625	1.531	1.606	1.731	1.766	1.652	E5	5.80
12)	N-OCTACOSANE	1.620	1.550	1.604	1.701	1.787	1.653	E5	5.60
13)	N-TRIACONTANE	1.637	1.597	1.620	1.716	1.842	1.682	E5	5.94
14)	N-DOTRIACONTANE	1.660	1.629	1.646	1.766	1.889	1.718	E5	6.37
15)	N-TETRATRIACONTANE	1.725	1.679	1.727	1.874	1.957	1.792	E5	6.57
16)	N-HEXATRIACONTANE	1.701	1.646	1.702	1.853	1.896	1.760	E5	6.15
17)	N-OCTATRIACONTANE	1.590	1.534	1.594	1.742	1.753	1.643	E5	6.01
18)	N-TETRACONTANE	1.624	1.567	1.620	1.760	1.732	1.660	E5	4.92

(#) = Out of Range