

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE050819\
 Data File : FE030023.D
 Signal(s) : FID1B.CH
 Acq On : 08 May 2019 9:32
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: Sample.e
 Quant Time: May 08 11:41:33 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE041019.M
 Quant Title :
 QLast Update : Wed Apr 10 08:20:21 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	14.656	8838175	43.566 ug/ml
Target Compounds				
2)	N-DECANE	4.384	10203119	46.431 ug/ml
3)	N-DODECANE	6.396	10399216	46.426 ug/ml
4)	N-TETRADECANE	8.192	10555907	45.710 ug/ml
5)	N-HEXADECANE	9.796	10644714	45.274 ug/ml
6)	N-OCTADECANE	11.241	10655315	44.875 ug/ml
7)	N-EICOSANE	12.554	10524067	44.416 ug/ml
8)	N-DOCOSANE	13.755	10347096	44.060 ug/ml
10)	N-TETRACOSANE	14.863	10123541	43.621 ug/ml
11)	N-HEXACOSANE	15.887	9851448	42.721 ug/ml
12)	N-OCTACOSANE	16.839	9631018	41.827 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE050819\
Data File : FE030023.D
Signal(s) : FID1B.CH
Acq On : 08 May 2019 9:32
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: Sample.e
Quant Time: May 08 11:41:33 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE041019.M
Quant Title :
QLast Update : Wed Apr 10 08:20:21 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

