

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE051322\
 Data File : FE035676.D
 Signal(s) : FID1B.CH
 Acq On : 14 May 2022 4:21
 Operator : YP\AJ
 Sample : 50 TRPH STD
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: May 14 04:52:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE051322.M
 Quant Title :
 QLast Update : Fri May 13 15:02:22 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.047	5382478	47.940 ug/ml
Target Compounds			
2) N-DECANE	3.825	6249609	49.131 ug/ml
3) N-DODECANE	5.820	6234043	48.750 ug/ml
4) N-TETRADECANE	7.610	6271509	48.575 ug/ml
5) N-HEXADECANE	9.208	6282227	48.617 ug/ml
6) N-OCTADECANE	10.647	6318777	48.450 ug/ml
7) N-EICOSANE	11.954	6208554	48.372 ug/ml
8) N-DOCOSANE	13.150	6158786	48.124 ug/ml
10) N-TETRACOSANE	14.251	6050544	47.742 ug/ml
11) N-HEXACOSANE	15.270	5923072	47.113 ug/ml
12) N-OCTACOSANE	16.216	5748144	46.532 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE051322\
 Data File : FE035676.D
 Signal(s) : FID1B.CH
 Acq On : 14 May 2022 4:21
 Operator : YP\AJ
 Sample : 50 TRPH STD
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: May 14 04:52:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE051322.M
 Quant Title :
 QLast Update : Fri May 13 15:02:22 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

