

rteres

Instrument :
FID_E
LabSampleId :
50 PPM TRPH STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 05/18/2022
Supervised By :mohammad ahmed 05/18/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE05162
Data File : FE035712.D
Signal(s) : FID1B.CH
Acq On : 17 May 2022 4:23
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE051322.M
Title :

Signal : FID1B.CH

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.822	3.797	3.862	BB	730741	6692436	98.50%	9.389%
2	5.815	5.787	5.855	BV	673337	6706136	98.70%	9.409%
3	7.604	7.563	7.643	BV	645539	6761414	99.52%	9.486%
4	9.201	9.169	9.233	BV	615842	6751025	99.37%	9.472%
5	10.640	10.583	10.678	BV	586062	6794155	100.00%	9.532%
6	11.947	11.903	11.993	BB	547903	6622030	97.47%	9.291%
7	13.143	13.096	13.189	BB	528775	6518324	95.94%	9.145%
8	14.039	13.988	14.108	BB	417870	5712004	84.07%	8.014%
9	14.243	14.196	14.281	BB	505033	6374757	93.83%	8.944%
10	15.261	15.210	15.306	BB	473120	6237322	91.80%	8.751%
11	16.209	16.141	16.261	BB	459222	6106222	89.87%	8.567%
Sum of corrected areas:						71275825		

FE051322.M Tue May 17 13:33:29 2022