

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE051721\
 Data File : FE033810.D
 Signal(s) : FID1B.CH
 Acq On : 17 May 2021 13:39
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: May 17 15:42:04 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE050521.M
 Quant Title :
 QLast Update : Thu May 06 02:18:38 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.726	21353947	52.833 ug/ml
Target Compounds			
2) N-DECANE	4.466	23569106	54.938 ug/ml
3) N-DODECANE	6.478	23931805	54.691 ug/ml
4) N-TETRADECANE	8.275	24339595	54.032 ug/ml
5) N-HEXADECANE	9.876	24091516	53.459 ug/ml
6) N-OCTADECANE	11.319	24555010	53.307 ug/ml
7) N-EICOSANE	12.629	24366999	53.081 ug/ml
8) N-DOCOSANE	13.827	24341078	52.790 ug/ml
10) N-TETRACOSANE	14.933	24277733	52.580 ug/ml
11) N-HEXACOSANE	15.953	23926940	52.147 ug/ml
12) N-OCTACOSANE	16.904	23745895	51.774 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE051721\
Data File : FE033810.D
Signal(s) : FID1B.CH
Acq On : 17 May 2021 13:39
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: May 17 15:42:04 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE050521.M
Quant Title :
QLast Update : Thu May 06 02:18:38 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

