

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE051820\
 Data File : FE032037.D
 Signal(s) : FID1B.CH
 Acq On : 18 May 2020 19:04
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: May 19 00:48:15 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE050720.M
 Quant Title :
 QLast Update : Thu May 07 13:58:31 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.366	19954168	51.482 ug/ml
Target Compounds			
2) N-DECANE	4.134	21811814	54.714 ug/ml
3) N-DODECANE	6.133	21700794	53.688 ug/ml
4) N-TETRADECANE	7.924	23279339	55.742 ug/ml
5) N-HEXADECANE	9.521	22908509	53.548 ug/ml
6) N-OCTADECANE	10.962	23674225	52.776 ug/ml
7) N-EICOSANE	12.270	23143544	52.187 ug/ml
8) N-DOCOSANE	13.469	23297443	51.981 ug/ml
10) N-TETRACOSANE	14.573	23095027	52.175 ug/ml
11) N-HEXACOSANE	15.591	22284550	50.933 ug/ml
12) N-OCTACOSANE	16.543	21873756	50.371 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE051820\
Data File : FE032037.D
Signal(s) : FID1B.CH
Acq On : 18 May 2020 19:04
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: May 19 00:48:15 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE050720.M
Quant Title :
QLast Update : Thu May 07 13:58:31 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

