

rteres

Instrument :
FID_E
LabSampleId :
50 PPM TRPH STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 05/19/2022
Supervised By :mohammad ahmed 05/19/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE05182
Data File : FE035735.D
Signal(s) : FID1B.CH
Acq On : 19 May 2022 1:04
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE051322.M
Title :

Signal : FID1B.CH

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.827	3.797	3.855	BV	802283	7364335	100.00%	9.669%
2	5.821	5.790	5.867	BB	743476	7335745	99.61%	9.631%
3	7.611	7.577	7.644	BV	711631	7364568	100.00%	9.669%
4	9.209	9.168	9.264	BV	632583	7332077	99.56%	9.626%
5	10.649	10.589	10.682	BB	654178	7312968	99.30%	9.601%
6	11.956	11.912	12.004	BB	598230	7056543	95.82%	9.265%
7	13.151	13.085	13.197	BB	555041	6887450	93.52%	9.043%
8	14.048	14.004	14.082	BV	414594	5853889	79.49%	7.686%
9	14.254	14.179	14.290	BB	510778	6699524	90.97%	8.796%
10	15.271	15.219	15.311	BB	491151	6535816	88.75%	8.581%
11	16.220	16.150	16.258	BB	454104	6423707	87.22%	8.434%
Sum of corrected areas:						76166622		

FE051322.M Thu May 19 12:05:33 2022