

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE051923AL\
 Data File : FE041987.D
 Signal(s) : FID1B.ch
 Acq On : 18 May 2023 11:26
 Operator : YP\AJ
 Sample : 02811-02DL 10X
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 72-11995-EPH-2DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/19/2023
 Supervised By :Ankita Jodhani 05/19/2023

Integration File: autoint1.e
 Quant Time: May 19 01:40:45 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 051823.M
 Quant Title : GC Extractables
 QLast Update : Thu May 18 06:27:12 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.965	709094	4.016 ug/mlm
Spiked Amount	50.000	Recovery =	8.03%
12) S 1-chlorooctadecane (S...	13.402	555368	4.180 ug/ml
Spiked Amount	50.000	Recovery =	8.36%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE051923AL\
Data File : FE041987.D
Signal(s) : FID1B.ch
Acq On : 18 May 2023 11:26
Operator : YP\AJ
Sample : 02811-02DL 10X
Misc :
ALS Vial : 15 Sample Multiplier: 1


Instrument :
FID_E
ClientSampleId :
72-11995-EPH-2DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/19/2023
Supervised By :Ankita Jodhani 05/19/2023

Integration File: autoint1.e
Quant Time: May 19 01:40:45 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 051823.
Quant Title : GC Extractables
QLast Update : Thu May 18 06:27:12 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.965 min
Delta R.T.: -0.003 min
Response: 709094
Conc: 4.02 ug/ml

Instrument :
FID_E
ClientSampleId :
72-11995-EPH-2DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/19/2023
Supervised By :Ankita Jodhani 05/19/2023



#12 1-chlorooctadecane (SURR)

R.T.: 13.402 min
Delta R.T.: -0.002 min
Response: 555368
Conc: 4.18 ug/ml