

rteres

Instrument :
FID_E
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE05232
Data File : FE042173.D
Signal(s) : FID1B.ch
Acq On : 24 May 2023 08:02
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 05/24/2023
Supervised By :Ankita Jodhani 05/24/2023

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 051823.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.346	3.307	3.385	BB	326006	3244742	81.70%	4.596%
2	4.596	4.555	4.637	BB	329124	3319383	83.58%	4.702%
3	6.302	6.264	6.346	BB	383517	3685962	92.81%	5.221%
4	6.764	6.727	6.799	BB	337572	3372546	84.92%	4.777%
5	7.400	7.358	7.446	BB	359498	3563446	89.72%	5.047%
6	8.592	8.553	8.630	BB	320907	3351283	84.38%	4.747%
7	10.200	10.156	10.243	BB	317470	3479817	87.62%	4.929%
8	11.644	11.604	11.690	BB	309396	3516461	88.54%	4.981%
9	11.963	11.916	12.014	BB	363662	3971579	100.00%	5.625%
10	12.952	12.907	12.996	BB	284355	3365567	84.74%	4.767%
11	13.399	13.348	13.443	BB	247334	2919297	73.50%	4.135%
12	13.565	13.518	13.608	BB	280233	3290202	82.84%	4.660%
13	14.151	14.103	14.193	BB	268514	3243789	81.68%	4.594%
14	15.254	15.202	15.301	BB	255148	3192155	80.37%	4.521%
15	16.275	16.237	16.322	BB	234191	3078285	77.51%	4.360%
16	17.225	17.177	17.267	BB	221896	2984093	75.14%	4.227%
17	18.113	18.056	18.153	BB	222013	3046253	76.70%	4.315%
18	18.943	18.891	18.984	BB	207672	2984022	75.13%	4.227%
19	19.728	19.676	19.766	BB	196820	2828150	71.21%	4.006%
20	20.468	20.301	20.503	BB	189854	2923624	73.61%	4.141%
21	21.217	21.161	21.269	BB	151460	2667903	67.17%	3.779%
22	22.152	22.087	22.211	BB	108583	2574018	64.81%	3.646%
Sum of corrected areas:							70602575	

Aliphatic EPH 051823.M Wed May 24 08:29:02 2023