

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE052820\
 Data File : FE032087.D
 Signal(s) : FID1B.CH
 Acq On : 28 May 2020 11:41
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: May 29 01:00:33 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE050720.M
 Quant Title :
 QLast Update : Thu May 07 13:58:31 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.367	19011416	49.050 ug/ml
Target Compounds			
1) N-OCTANE	2.129	20354202	50.291 ug/ml
2) N-DECANE	4.130	20834220	52.262 ug/ml
3) N-DODECANE	6.130	20476605	50.659 ug/ml
4) N-TETRADECANE	7.921	22236286	53.244 ug/ml
5) N-HEXADECANE	9.519	21905216	51.203 ug/ml
6) N-OCTADECANE	10.960	22637326	50.464 ug/ml
7) N-EICOSANE	12.269	22152020	49.951 ug/ml
8) N-DOCOSANE	13.468	22192067	49.514 ug/ml
10) N-TETRACOSANE	14.574	21956049	49.602 ug/ml
11) N-HEXACOSANE	15.594	21374634	48.853 ug/ml
12) N-OCTACOSANE	16.544	20994107	48.346 ug/ml
13) N-TRIACONTANE	17.434	20831576	47.940 ug/ml
14) N-DOTRIACONTANE	18.268	20012515	47.988 ug/ml
15) N-TETRATRIACONTANE	19.053	19359738	48.978 ug/ml
16) N-HEXATRIACONTANE	19.793	18264751	49.147 ug/ml
17) N-OCTATRIACONTANE	20.496	17098669	50.858 ug/ml
18) N-TETRACONTANE	21.303	17192003	50.811 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE052820\
Data File : FE032087.D
Signal(s) : FID1B.CH
Acq On : 28 May 2020 11:41
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: May 29 01:00:33 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE050720.M
Quant Title :
QLast Update : Thu May 07 13:58:31 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

