

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE060220\
 Data File : FE032113.D
 Signal(s) : FID1B.CH
 Acq On : 02 Jun 2020 13:20
 Operator : DD\AJ
 Sample : PB129236BS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jun 03 01:59:44 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE050720.M
 Quant Title :
 QLast Update : Thu May 07 13:58:31 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.347	8108324	20.920 ug/ml
Target Compounds			
2) N-DECANE	4.119	8591138	21.551 ug/ml
3) N-DODECANE	6.116	8882118	21.974 ug/mlm
4) N-TETRADECANE	7.904	9435001	22.592 ug/mlm
5) N-HEXADECANE	9.502	9536357	22.291 ug/ml
6) N-OCTADECANE	10.942	9901408	22.073 ug/ml
7) N-EICOSANE	12.251	9679106	21.825 ug/ml
8) N-DOCOSANE	13.449	9666483	21.568 ug/ml
10) N-TETRACOSANE	14.553	9537449	21.547 ug/ml
11) N-HEXACOSANE	15.574	9103213	20.806 ug/ml
12) N-OCTACOSANE	16.525	8808343	20.284 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE060220\
Data File : FE032113.D
Signal(s) : FID1B.CH
Acq On : 02 Jun 2020 13:20
Operator : DD\AJ
Sample : PB129236BS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jun 03 01:59:44 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE050720.M
Quant Title :
QLast Update : Thu May 07 13:58:31 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

