

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE060223AL\
 Data File : FE042368.D
 Signal(s) : FID1B.ch
 Acq On : 02 Jun 2023 11:22
 Operator : YP\AJ
 Sample : 02988-02DL 10X
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 ORA-1644-EPH-2DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/05/2023
 Supervised By :Ankita Jodhani 06/05/2023

Integration File: autoint1.e
 Quant Time: Jun 05 01:15:18 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 051823.M
 Quant Title : GC Extractables
 QLast Update : Thu May 18 06:27:12 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.948	486559	2.755 ug/mlm
Spiked Amount	50.000	Recovery	= 5.51%
12) S 1-chlorooctadecane (S...	13.385	424275	3.193 ug/mlm
Spiked Amount	50.000	Recovery	= 6.39%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE060223AL\
Data File : FE042368.D
Signal(s) : FID1B.ch
Acq On : 02 Jun 2023 11:22
Operator : YP\AJ
Sample : 02988-02DL 10X
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :

FID_E

ClientSampleId :

ORA-1644-EPH-2DL

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 06/05/2023

Supervised By :Ankita Jodhani 06/05/2023

Integration File: autoint1.e

Quant Time: Jun 05 01:15:18 2023

Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 051823.

Quant Title : GC Extractables

QLast Update : Thu May 18 06:27:12 2023


Response via : Initial Calibration


Integrator: ChemStation

Volume Inj. : 1 ul

Signal Phase : Rxi-1ms

Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.948 min
Delta R.T.: -0.020 min
Response: 486559
Conc: 2.76 ug/ml

Instrument : FID_E
ClientSampleId :
ORA-1644-EPH-2DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/05/2023
Supervised By :Ankita Jodhani 06/05/2023



#12 1-chlorooctadecane (SURR)

R.T.: 13.385 min
Delta R.T.: -0.019 min
Response: 424275
Conc: 3.19 ug/ml m