

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE060321\
 Data File : FE033873.D
 Signal(s) : FID1B.CH
 Acq On : 03 Jun 2021 16:07
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM TRPH STD

Integration File: Sample.e
 Quant Time: Jun 04 03:29:35 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE050521.M
 Quant Title :
 QLast Update : Thu May 06 02:18:38 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.731	19483696	48.206 ug/ml
Target Compounds			
1) N-OCTANE	2.409	21196313	51.101 ug/ml
2) N-DECANE	4.467	22662936	52.826 ug/ml
3) N-DODECANE	6.480	23069899	52.721 ug/ml
4) N-TETRADECANE	8.277	23432524	52.019 ug/ml
5) N-HEXADECANE	9.879	23165809	51.405 ug/ml
6) N-OCTADECANE	11.322	23470520	50.952 ug/ml
7) N-EICOSANE	12.633	23030048	50.169 ug/ml
8) N-DOCOSANE	13.833	22650719	49.124 ug/ml
10) N-TETRACOSANE	14.939	22193120	48.066 ug/ml
11) N-HEXACOSANE	15.961	21235004	46.280 ug/ml
12) N-OCTACOSANE	16.913	20445347	44.577 ug/ml
13) N-TRIACONTANE	17.801	19691555	43.196 ug/ml
14) N-DOTRIACONTANE	18.634	18708805	42.030 ug/ml
15) N-TETRATRIACONTANE	19.419	18103462	42.115 ug/ml
16) N-HEXATRIACONTANE	20.160	17259137	43.049 ug/ml
17) N-OCTATRIACONTANE	20.906	16619521	45.096 ug/ml
18) N-TETRACONTANE	21.838	17134764	47.386 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE060321\
Data File : FE033873.D
Signal(s) : FID1B.CH
Acq On : 03 Jun 2021 16:07
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
50 PPM TRPH STD

Integration File: Sample.e
Quant Time: Jun 04 03:29:35 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE050521.M
Quant Title :
QLast Update : Thu May 06 02:18:38 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

