

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_E\Data\FE060425AL\  
 Data File : FE054175.D  
 Signal(s) : FID1B.ch  
 Acq On : 04 Jun 2025 17:53  
 Operator : YP\AJ  
 Sample : Q2194-01  
 Misc :  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 FID\_E  
 ClientSampleId :  
 COMP-12

Manual Integrations  
 APPROVED

Reviewed By :Yogesh Patel 06/05/2025  
 Supervised By :mohammad ahmed 06/06/2025

Integration File: autoint1.e  
 Quant Time: Jun 05 06:27:16 2025  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_E\methods\Aliphatic EPH 051425.M  
 Quant Title : GC Extractables  
 QLast Update : Wed May 14 12:15:41 2025  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 ul  
 Signal Phase : Rxi-1ms  
 Signal Info : 20M x 0.18mm x 0.18um

| Compound                       | R.T.   | Response   | Conc Units    |
|--------------------------------|--------|------------|---------------|
| -----                          |        |            |               |
| System Monitoring Compounds    |        |            |               |
| 9) S ortho-Terphenyl (SURR)    | 11.846 | 12039843   | 74.086 ug/mlm |
| Spiked Amount 50.000           |        | Recovery = | 148.17%       |
| 12) S 1-chlorooctadecane (S... | 13.289 | 9387618    | 79.187 ug/mlm |
| Spiked Amount 50.000           |        | Recovery = | 158.37%       |

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_E\Data\FE060425AL\  
Data File : FE054175.D  
Signal(s) : FID1B.ch  
Acq On : 04 Jun 2025 17:53  
Operator : YP\AJ  
Sample : Q2194-01  
Misc :  
ALS Vial : 10 Sample Multiplier: 1


Instrument :  
FID\_E  
ClientSampleId :  
COMP-12


Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 06/05/2025  
Supervised By :mohammad ahmed 06/06/2025

Integration File: autoint1.e  
Quant Time: Jun 05 06:27:16 2025  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_E\methods\Aliphatic EPH 051425.  
Quant Title : GC Extractables  
QLast Update : Wed May 14 12:15:41 2025  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.846 min  
Delta R.T.: -0.003 min  
Response: 12039843  
Conc: 74.09 ug/ml

Instrument :  
FID\_E  
ClientSampleId :  
COMP-12

Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 06/05/2025  
Supervised By :mohammad ahmed 06/06/2025



#12 1-chlorooctadecane (SURR)

R.T.: 13.289 min  
Delta R.T.: -0.006 min  
Response: 9387618  
Conc: 79.19 ug/ml m