

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE060521\
 Data File : FE033897.D
 Signal(s) : FID1B.CH
 Acq On : 05 Jun 2021 18:00
 Operator : DD\AJ
 Sample : 100 TRPH STD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 100 TRPH STD

Integration File: autoint1.e
 Quant Time: Jun 07 02:17:43 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE060521.M
 Quant Title :
 QLast Update : Mon Jun 07 02:17:16 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.730	36298792	98.719 ug/ml
Target Compounds			
1) N-OCTANE	2.399	37704681	98.141 ug/ml
2) N-DECANE	4.463	39071194	99.191 ug/ml
3) N-DODECANE	6.479	39913018	99.429 ug/ml
4) N-TETRADECANE	8.277	40694450	98.864 ug/ml
5) N-HEXADECANE	9.879	40444199	98.536 ug/ml
6) N-OCTADECANE	11.322	41352377	98.396 ug/ml
7) N-EICOSANE	12.633	41196649	98.477 ug/ml
8) N-DOCOSANE	13.832	41387311	98.523 ug/ml
10) N-TETRACOSANE	14.939	41516750	98.713 ug/ml
11) N-HEXACOSANE	15.960	41278861	98.863 ug/ml
12) N-OCTACOSANE	16.911	41336160	98.923 ug/ml
13) N-TRIACONTANE	17.800	41160505	98.758 ug/ml
14) N-DOTRIACONTANE	18.634	40416240	98.724 ug/ml
15) N-TETRATRIACONTANE	19.418	39737759	98.995 ug/ml
16) N-HEXATRIACONTANE	20.158	38015833	99.242 ug/ml
17) N-OCTATRIACONTANE	20.905	35631798	99.135 ug/ml
18) N-TETRACONTANE	21.838	35365183	99.160 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE060521\
Data File : FE033897.D
Signal(s) : FID1B.CH
Acq On : 05 Jun 2021 18:00
Operator : DD\AJ
Sample : 100 TRPH STD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
100 TRPH STD

Integration File: autoint1.e
Quant Time: Jun 07 02:17:43 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE060521.M
Quant Title :
QLast Update : Mon Jun 07 02:17:16 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

