

rteres

Instrument :
FID_E
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/06/2025
Supervised By :mohammad ahmed 06/09/2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE06052
Data File : FE054193.D
Signal(s) : FID1B.ch
Acq On : 05 Jun 2025 15:42
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 051425.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.184	3.124	3.315	BB	209682	2200934	73.32%	4.171%
2	4.447	4.390	4.549	BV	213905	2265565	75.48%	4.294%
3	6.162	6.112	6.315	BV	196295	2505454	83.47%	4.748%
4	6.628	6.564	6.784	BB	223506	2375171	79.13%	4.501%
5	7.264	7.230	7.392	BV	181758	2475139	82.46%	4.691%
6	8.463	8.325	8.605	BB	208644	2408218	80.23%	4.564%
7	10.075	9.955	10.229	BB	207992	2470117	82.29%	4.681%
8	11.521	11.490	11.627	BB	210892	2491530	83.00%	4.722%
9	11.835	11.777	11.905	BV	263115	3001731	100.00%	5.689%
10	12.835	12.797	12.912	PV	202707	2469034	82.25%	4.679%
11	13.281	13.214	13.372	BB	188089	2255273	75.13%	4.274%
12	13.448	13.390	13.534	BB	201906	2440422	81.30%	4.625%
13	14.036	14.005	14.124	VB	199252	2450957	81.65%	4.645%
14	15.142	15.040	15.214	BB	190942	2487878	82.88%	4.715%
15	16.164	16.100	16.265	BB	187581	2469129	82.26%	4.679%
16	17.116	17.039	17.189	BB	179722	2468829	82.25%	4.679%
17	18.005	17.915	18.090	PB	180740	2553316	85.06%	4.839%
18	18.839	18.800	18.910	BB	175832	2557671	85.21%	4.847%
19	19.624	19.590	19.692	BV	157683	2347156	78.19%	4.448%
20	20.366	20.320	20.449	BV	137092	2202109	73.36%	4.173%
21	21.094	21.040	21.185	BB	106605	1967393	65.54%	3.729%
22	21.986	21.919	22.127	PB	73598	1903137	63.40%	3.607%
Sum of corrected areas:						52766163		

Aliphatic EPH 051425.M Fri Jun 06 02:07:19 2025