

rteres

Instrument :
FID_E
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/07/2022
Supervised By :Ankita Jodhani 06/07/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE06062
Data File : FE035914.D
Signal(s) : FID1B.CH
Acq On : 06 Jun 2022 15:46
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 051022.M
Title : GC Extractables

Signal : FID1B.CH

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.802	2.773	2.840	BB	295171	2361001	90.87%	4.722%
2	3.817	3.784	3.856	BB	263947	2346935	90.33%	4.694%
3	5.385	5.359	5.423	BB	284306	2593523	99.82%	5.187%
4	5.807	5.774	5.854	BB	245645	2355959	90.68%	4.712%
5	6.434	6.390	6.475	BB	276349	2598231	100.00%	5.196%
6	7.596	7.552	7.629	BB	228369	2333705	89.82%	4.667%
7	9.193	9.149	9.238	BB	212405	2323389	89.42%	4.646%
8	10.632	10.589	10.673	BB	206782	2326443	89.54%	4.653%
9	10.953	10.907	10.999	BB	233934	2556836	98.41%	5.113%
10	11.938	11.895	11.988	BB	194982	2295758	88.36%	4.591%
11	12.383	12.338	12.426	BB	170506	1970263	75.83%	3.940%
12	12.549	12.504	12.588	BB	189424	2296494	88.39%	4.593%
13	13.134	13.099	13.183	BB	186348	2294590	88.31%	4.589%
14	14.234	14.189	14.288	BB	179008	2290834	88.17%	4.581%
15	15.253	15.207	15.296	BB	171335	2265531	87.20%	4.531%
16	16.200	16.132	16.239	BB	168041	2267959	87.29%	4.536%
17	17.085	17.014	17.125	BB	163449	2265722	87.20%	4.531%
18	17.917	17.860	17.967	BB	148362	2187152	84.18%	4.374%
19	18.699	18.635	18.734	BV	144737	2070877	79.70%	4.141%
20	19.437	19.400	19.472	M	132163	1981636	76.27%	3.963%
21	20.172	20.125	20.215	M	108209	1941081	74.71%	3.882%
22	21.088	20.979	21.184	BB	82723	2079662	80.04%	4.159%
Sum of corrected areas:							50003581	

Aliphatic EPH 051022.M Tue Jun 07 13:22:13 2022