

Instrument :
FID_E
LabSampleId :
20 PPM ALIPHATIC HC STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/07/2023
Supervised By :Ankita Jodhani 06/07/2023

rteres

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE
Data File : FE042441.D
Signal(s) : FID1B.ch
Acq On : 07 Jun 2023 00:37
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 051823.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.331	3.301	3.376	BB	264061	2681472	83.32%	4.406%
2	4.584	4.554	4.625	BB	276496	2768912	86.04%	4.549%
3	6.293	6.266	6.343	BB	295676	2979353	92.58%	4.895%
4	6.752	6.716	6.791	BB	276014	2789839	86.69%	4.584%
5	7.391	7.349	7.444	BB	289688	2860240	88.88%	4.699%
6	8.582	8.546	8.623	BB	263359	2771779	86.13%	4.554%
7	10.190	10.154	10.235	BB	256945	2873801	89.30%	4.722%
8	11.634	11.599	11.678	BB	251641	2896875	90.01%	4.760%
9	11.954	11.916	11.996	BB	298221	3218255	100.00%	5.288%
10	12.945	12.908	12.987	BB	239479	2768146	86.01%	4.548%
11	13.390	13.357	13.434	BB	198557	2367825	73.57%	3.890%
12	13.557	13.519	13.599	BB	215545	2696566	83.79%	4.430%
13	14.144	14.068	14.186	BB	216803	2699782	83.89%	4.436%
14	15.248	15.207	15.294	BB	214511	2640188	82.04%	4.338%
15	16.269	16.221	16.316	BB	202257	2627655	81.65%	4.317%
16	17.219	17.175	17.259	BB	203279	2672324	83.04%	4.391%
17	18.108	18.061	18.156	BB	209417	2932968	91.14%	4.819%
18	18.940	18.889	18.995	BB	211167	2991255	92.95%	4.915%
19	19.726	19.676	19.762	BB	193397	2765267	85.92%	4.543%
20	20.467	20.392	20.529	VB	173420	2752149	85.52%	4.522%
21	21.216	21.161	21.276	BB	142127	2624111	81.54%	4.311%
22	22.151	22.087	22.227	BB	100916	2486169	77.25%	4.085%
Sum of corrected areas:						60864929		

Aliphatic EPH 051823.M Wed Jun 07 05:53:57 2023