

rteres

Instrument :
FID_E
LabSampleId :
20 PPM ALIPHATIC HC STD

Area Percent

Manual IntegrationsAPPROVED

Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Da
Data File : FE048583.D
Signal(s) : FID1B.ch
Acq On : 06 Jun 2024 07:03
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 6 Sample Multiplier: 1

Reviewed By :Yogesh Patel 06/07/2024
Supervised By :Ankita Jodhani 06/07/2024

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH
052524.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.333	3.298	3.375	BB	277963	2698078	90.67%	4.654%
2	4.582	4.549	4.632	BB	295068	2797183	94.00%	4.825%
3	6.287	6.250	6.372	BB	276277	2742834	92.17%	4.731%
4	6.747	6.709	6.795	BB	281846	2743220	92.19%	4.732%
5	7.384	7.339	7.476	BB	252233	2570623	86.39%	4.434%
6	8.575	8.534	8.633	BB	253205	2632882	88.48%	4.541%
7	10.183	10.142	10.247	BB	244487	2631959	88.45%	4.540%
8	11.626	11.581	11.681	BB	232214	2600969	87.41%	4.486%
9	11.946	11.890	11.998	BB	254614	2689658	90.39%	4.639%
10	12.934	12.890	12.992	BB	204686	2480667	83.36%	4.279%
11	13.380	13.340	13.437	BB	173169	2022121	67.95%	3.488%
12	13.547	13.498	13.604	BB	201622	2418272	81.27%	4.171%
13	14.132	14.101	14.193	VB	202613	2445976	82.20%	4.219%
14	15.235	15.160	15.299	BB	194746	2551415	85.74%	4.401%
15	16.254	16.207	16.323	BB	204535	2594741	87.20%	4.476%
16	17.204	17.066	17.272	BB	201906	2720818	91.43%	4.693%
17	18.091	18.041	18.151	BB	218541	2911748	97.85%	5.022%
18	18.922	18.855	18.986	BB	217144	2975741	100.00%	5.133%
19	19.705	19.648	19.781	BB	198878	2863917	96.24%	4.940%
20	20.444	20.401	20.513	BB	188944	2705823	90.93%	4.667%
21	21.182	21.121	21.263	BB	145766	2627918	88.31%	4.533%
22	22.100	22.040	22.229	BV	105416	2549208	85.67%	4.397%
Sum of corrected areas:						57975770		

Aliphatic EPH 052524.M Fri Jun 07 01:50:04 2024

Page 1