

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE061021\
 Data File : FE033954.D
 Signal(s) : FID1B.CH
 Acq On : 10 Jun 2021 9:12
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jun 11 02:03:32 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE060521.M
 Quant Title :
 QLast Update : Mon Jun 07 02:30:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	14.713	19054138	49.788 ug/ml
Target Compounds				
2)	N-DECANE	4.453	21643572	53.067 ug/ml
3)	N-DODECANE	6.466	22029160	53.299 ug/ml
4)	N-TETRADECANE	8.262	22337099	52.459 ug/ml
5)	N-HEXADECANE	9.863	22019469	51.847 ug/ml
6)	N-OCTADECANE	11.305	22356777	51.257 ug/ml
7)	N-EICOSANE	12.616	22071252	50.800 ug/ml
8)	N-DOCOSANE	13.814	21943972	50.277 ug/ml
10)	N-TETRACOSANE	14.919	21728597	49.623 ug/ml
11)	N-HEXACOSANE	15.940	21235773	48.892 ug/ml
12)	N-OCTACOSANE	16.890	20886193	48.001 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE061021\
Data File : FE033954.D
Signal(s) : FID1B.CH
Acq On : 10 Jun 2021 9:12
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jun 11 02:03:32 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE060521.M
Quant Title :
QLast Update : Mon Jun 07 02:30:45 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

