

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE061021\
 Data File : FE033961.D
 Signal(s) : FID1B.CH
 Acq On : 10 Jun 2021 16:34
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jun 11 01:49:35 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE060521.M
 Quant Title :
 QLast Update : Mon Jun 07 02:30:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.711	18489719	48.313 ug/ml
Target Compounds			
2) N-DECANE	4.451	21053652	51.621 ug/ml
3) N-DODECANE	6.464	21410534	51.802 ug/ml
4) N-TETRADECANE	8.260	21704254	50.973 ug/ml
5) N-HEXADECANE	9.862	21403276	50.396 ug/ml
6) N-OCTADECANE	11.304	21722780	49.804 ug/ml
7) N-EICOSANE	12.614	21440096	49.348 ug/ml
8) N-DOCOSANE	13.813	21310193	48.824 ug/ml
10) N-TETRACOSANE	14.918	21083049	48.148 ug/ml
11) N-HEXACOSANE	15.939	20588070	47.401 ug/ml
12) N-OCTACOSANE	16.888	20234933	46.505 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE061021\
Data File : FE033961.D
Signal(s) : FID1B.CH
Acq On : 10 Jun 2021 16:34
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jun 11 01:49:35 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE060521.M
Quant Title :
QLast Update : Mon Jun 07 02:30:45 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

