

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE061321\
 Data File : FE033974.D
 Signal(s) : FID1B.CH
 Acq On : 13 Jun 2021 14:33
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM TRPH STD

Integration File: Sample.e
 Quant Time: Jun 14 01:11:44 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE060521.M
 Quant Title :
 QLast Update : Mon Jun 07 02:30:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.712	19312940	50.464 ug/ml
Target Compounds			
2) N-DECANE	4.452	21700111	53.206 ug/ml
3) N-DODECANE	6.465	22051591	53.353 ug/ml
4) N-TETRADECANE	8.261	22348755	52.486 ug/ml
5) N-HEXADECANE	9.862	22094646	52.024 ug/ml
6) N-OCTADECANE	11.304	22473098	51.524 ug/ml
7) N-EICOSANE	12.615	22221411	51.146 ug/ml
8) N-DOCOSANE	13.813	22113431	50.665 ug/ml
10) N-TETRACOSANE	14.918	22103460	50.479 ug/ml
11) N-HEXACOSANE	15.939	21559892	49.638 ug/ml
12) N-OCTACOSANE	16.891	21362263	49.095 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE061321\
Data File : FE033974.D
Signal(s) : FID1B.CH
Acq On : 13 Jun 2021 14:33
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
50 PPM TRPH STD

Integration File: Sample.e
Quant Time: Jun 14 01:11:44 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE060521.M
Quant Title :
QLast Update : Mon Jun 07 02:30:45 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

