

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE061422\
 Data File : FE036025.D
 Signal(s) : FID1B.CH
 Acq On : 14 Jun 2022 13:11
 Operator : YP\AJ
 Sample : N1645-03
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jun 15 05:04:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE061022.M
 Quant Title :
 QLast Update : Sat Jun 11 00:12:39 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	13.965	2181055	20.262 ug/ml
Target Compounds			
2) N-DECANE	3.737	118427	0.973 ug/ml
3) N-DODECANE	5.728	116039	0.950 ug/ml
4) N-TETRADECANE	7.519	115380	0.935 ug/mlm
5) N-HEXADECANE	9.119	125485	1.023 ug/mlm
6) N-OCTADECANE	10.559	123357	0.998 ug/mlm
7) N-EICOSANE	11.866	132370	1.083 ug/ml
8) N-DOCOSANE	13.064	137242	1.124 ug/ml
10) N-TETRACOSANE	14.167	143126	1.182 ug/ml
11) N-HEXACOSANE	15.188	137843	1.146 ug/ml
12) N-OCTACOSANE	16.137	148013	1.248 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE061422\
Data File : FE036025.D
Signal(s) : FID1B.CH
Acq On : 14 Jun 2022 13:11
Operator : YP\AJ
Sample : N1645-03
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jun 15 05:04:46 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE061022.M
Quant Title :
QLast Update : Sat Jun 11 00:12:39 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

