

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE061423AL\
 Data File : FE042602.D
 Signal(s) : FID1B.ch
 Acq On : 14 Jun 2023 08:52
 Operator : YP\AJ
 Sample : 03154-04DL 5X
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 CHRT-26598-E2DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/15/2023
 Supervised By :Ankita Jodhani 06/15/2023

Integration File: autoint1.e
 Quant Time: Jun 15 01:24:52 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 051823.M
 Quant Title : GC Extractables
 QLast Update : Thu May 18 06:27:12 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.940	1539052	8.716 ug/mlm
Spiked Amount 50.000		Recovery =	17.43%
12) S 1-chlorooctadecane (S...	13.377	1018081	7.662 ug/mlm
Spiked Amount 50.000		Recovery =	15.32%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE061423AL\
Data File : FE042602.D
Signal(s) : FID1B.ch
Acq On : 14 Jun 2023 08:52
Operator : YP\AJ
Sample : 03154-04DL 5X
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :

FID_E

ClientSampleId :

CHRT-26598-E2DL

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 06/15/2023

Supervised By :Ankita Jodhani 06/15/2023

Integration File: autoint1.e

Quant Time: Jun 15 01:24:52 2023

Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 051823.

Quant Title : GC Extractables

QLast Update : Thu May 18 06:27:12 2023


Response via : Initial Calibration


Integrator: ChemStation

Volume Inj. : 1 ul

Signal Phase : Rxi-1ms

Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.940 min
Delta R.T.: -0.028 min
Response: 1539052
Conc: 8.72 ug/ml

Instrument :
FID_E
ClientSampleId :
CHRT-26598-E2DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/15/2023
Supervised By :Ankita Jodhani 06/15/2023



#12 1-chlorooctadecane (SURR)

R.T.: 13.377 min
Delta R.T.: -0.027 min
Response: 1018081
Conc: 7.66 ug/ml m