

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE061520\
 Data File : FE032175.D
 Signal(s) : FID1B.CH
 Acq On : 15 Jun 2020 20:34
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jun 16 01:05:19 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE061120.M
 Quant Title :
 QLast Update : Thu Jun 11 13:38:20 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.751	14976689	46.842 ug/ml
Target Compounds			
2) N-DECANE	4.499	18219180	49.023 ug/ml
3) N-DODECANE	6.514	18176815	48.768 ug/ml
4) N-TETRADECANE	8.308	17954160	48.482 ug/ml
5) N-HEXADECANE	9.909	17878502	48.244 ug/ml
6) N-OCTADECANE	11.350	18008877	48.046 ug/ml
7) N-EICOSANE	12.658	17734030	47.741 ug/ml
8) N-DOCOSANE	13.855	17486010	47.275 ug/ml
10) N-TETRACOSANE	14.958	17312253	46.640 ug/ml
11) N-HEXACOSANE	15.977	16787477	45.846 ug/ml
12) N-OCTACOSANE	16.924	16446483	44.999 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE061520\
Data File : FE032175.D
Signal(s) : FID1B.CH
Acq On : 15 Jun 2020 20:34
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jun 16 01:05:19 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE061120.M
Quant Title :
QLast Update : Thu Jun 11 13:38:20 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

