

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE061623AL\
 Data File : FE042671.D
 Signal(s) : FID1B.ch
 Acq On : 16 Jun 2023 21:17
 Operator : YP\AJ
 Sample : 03262-06
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Jun 17 01:08:56 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 051823.M
 Quant Title : GC Extractables
 QLast Update : Thu May 18 06:27:12 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.936	8760389	49.612 ug/ml
Spiked Amount 50.000		Recovery =	99.22%
12) S 1-chlorooctadecane (S...	13.372	6700039	50.422 ug/ml
Spiked Amount 50.000		Recovery =	100.84%
Target Compounds			
10) T n-Eicosane (C20)	12.987	273858	1.807 ug/ml
14) T n-Tetracosane (C24)	15.207f	193004	1.337 ug/ml
20) T n-Hexatriacontane (C36)	20.542f	261601	2.013 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE061623AL\
Data File : FE042671.D
Signal(s) : FID1B.ch
Acq On : 16 Jun 2023 21:17
Operator : YP\AJ
Sample : 03262-06
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jun 17 01:08:56 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 051823.M
Quant Title : GC Extractables
QLast Update : Thu May 18 06:27:12 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)


R.T.: 11.936 min
Delta R.T.: -0.032 min
Response: 8760389
Conc: 49.61 ug/ml

Instrument :
FID_E
ClientSampleId :




#10 n-Eicosane (C20)

R.T.: 12.987 min
Delta R.T.: 0.027 min
Response: 273858
Conc: 1.81 ug/ml




#12 1-chlorooctadecane (SURR)

R.T.: 13.372 min
Delta R.T.: -0.032 min
Response: 6700039
Conc: 50.42 ug/ml



#14 n-Tetracosane (C24)

R.T.: 15.207 min
Delta R.T.: -0.054 min
Response: 193004
Conc: 1.34 ug/ml



#20 n-Hexatriacontane (C36)

R.T.: 20.542 min
Delta R.T.: 0.065 min
Response: 261601
Conc: 2.01 ug/ml

Instrument :
FID_E
ClientSampleId :