

Instrument :
FID_E
LabSampleId :
20 PPM ALIPHATIC HC STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/19/2023
Supervised By :Ankita Jodhani 06/19/2023

rteres

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE
Data File : FE042681.D
Signal(s) : FID1B.ch
Acq On : 17 Jun 2023 02:53
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 051823.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.318	3.286	3.363	BB	283488	2876559	82.55%	4.656%
2	4.570	4.537	4.615	BB	290383	2927122	84.01%	4.738%
3	6.280	6.251	6.347	BB	316998	3173725	91.08%	5.137%
4	6.739	6.709	6.785	BB	294095	2966316	85.13%	4.802%
5	7.377	7.341	7.448	BB	290837	3057179	87.74%	4.949%
6	8.568	8.536	8.616	BB	286329	2944890	84.52%	4.767%
7	10.177	10.145	10.225	BB	264076	3055537	87.69%	4.946%
8	11.621	11.579	11.671	BB	265194	3101045	89.00%	5.020%
9	11.941	11.899	11.986	BB	315219	3484423	100.00%	5.640%
10	12.931	12.885	12.978	BB	254128	2991697	85.86%	4.843%
11	13.377	13.331	13.426	BB	213608	2598812	74.58%	4.207%
12	13.543	13.501	13.591	BB	247737	2934749	84.22%	4.751%
13	14.130	14.086	14.176	BB	229368	2903142	83.32%	4.699%
14	15.233	15.187	15.273	BB	225297	2831630	81.27%	4.584%
15	16.254	16.206	16.296	BB	206624	2707826	77.71%	4.383%
16	17.205	17.158	17.246	BB	188424	2563355	73.57%	4.149%
17	18.093	18.042	18.139	BB	185933	2537379	72.82%	4.107%
18	18.926	18.870	18.964	BB	169595	2448752	70.28%	3.964%
19	19.709	19.620	19.756	BB	163606	2387495	68.52%	3.865%
20	20.451	20.404	20.495	M	159045	2363343	67.83%	3.826%
21	21.198	21.141	21.248	BB	134158	2413466	69.26%	3.907%
22	22.127	22.053	22.180	BV	101191	2507801	71.97%	4.059%
Sum of corrected areas:						61776241		

Aliphatic EPH 051823.M Sat Jun 17 03:21:48 2023