

rteres

Instrument :
FID_E
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/17/2025
Supervised By :mohammad ahmed 06/18/2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE06162
Data File : FE054422.D
Signal(s) : FID1B.ch
Acq On : 17 Jun 2025 00:18
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 061025.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.175	3.090	3.298	BB	242431	2586717	75.08%	4.268%
2	4.438	4.383	4.575	BB	248698	2624830	76.19%	4.331%
3	6.153	6.103	6.393	BB	221182	2873425	83.41%	4.741%
4	6.620	6.567	6.785	BB	250983	2707549	78.59%	4.468%
5	7.255	7.193	7.385	BV	205377	2763030	80.20%	4.559%
6	8.454	8.387	8.618	BB	241707	2717416	78.88%	4.484%
7	10.066	10.020	10.225	BB	235699	2788772	80.95%	4.602%
8	11.513	11.470	11.637	BB	233279	2828344	82.10%	4.667%
9	11.826	11.758	11.943	BB	300512	3445127	100.00%	5.685%
10	12.826	12.720	12.897	BV	227468	2792445	81.05%	4.608%
11	13.271	13.230	13.403	BV	204124	2560725	74.33%	4.225%
12	13.439	13.403	13.545	VB	218764	2783137	80.78%	4.592%
13	14.026	13.992	14.112	VB	217846	2750415	79.83%	4.538%
14	15.131	15.032	15.233	BB	203749	2733553	79.35%	4.511%
15	16.154	16.110	16.233	BB	200172	2668852	77.47%	4.404%
16	17.106	17.070	17.182	BB	190837	2652197	76.98%	4.376%
17	17.995	17.960	18.057	BV	193368	2744413	79.66%	4.529%
18	18.829	18.790	18.887	BV	188756	2809489	81.55%	4.636%
19	19.614	19.580	19.682	BV	182212	2705665	78.54%	4.465%
20	20.355	20.320	20.412	BV	169770	2752948	79.91%	4.543%
21	21.082	21.040	21.150	BV	140647	2719822	78.95%	4.488%
22	21.970	21.920	22.107	BV	100105	2593254	75.27%	4.279%
Sum of corrected areas:						60602123		

Aliphatic EPH 061025.M Tue Jun 17 04:32:09 2025