

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062025AL\
 Data File : FE054521.D
 Signal(s) : FID1B.ch
 Acq On : 20 Jun 2025 15:51
 Operator : YP\AJ
 Sample : Q2362-01D
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 TP-11

Integration File: autoint1.e
 Quant Time: Jun 20 23:24:27 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 061025.M
 Quant Title : GC Extractables
 QLast Update : Tue Jun 10 07:24:04 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.824	6901123	41.765 ug/ml
Spiked Amount	50.000	Recovery	= 83.53%
12) S 1-chlorooctadecane (S...	13.268	5472467	44.875 ug/ml
Spiked Amount	50.000	Recovery	= 89.75%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062025AL\
Data File : FE054521.D
Signal(s) : FID1B.ch
Acq On : 20 Jun 2025 15:51
Operator : YP\AJ
Sample : Q2362-01D
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
TP-11

Integration File: autoint1.e
Quant Time: Jun 20 23:24:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 061025.M
Quant Title : GC Extractables
QLast Update : Tue Jun 10 07:24:04 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.824 min
Delta R.T.: -0.011 min
Response: 6901123
Conc: 41.77 ug/ml

Instrument :
FID_E
ClientSampleId :
TP-11



#12 1-chlorooctadecane (SURR)

R.T.: 13.268 min
Delta R.T.: -0.012 min
Response: 5472467
Conc: 44.88 ug/ml