

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062420\
 Data File : FE032247.D
 Signal(s) : FID1B.CH
 Acq On : 24 Jun 2020 19:22
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jun 25 01:14:38 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE061120.M
 Quant Title :
 QLast Update : Thu Jun 11 13:38:20 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.760	15224527	47.617 ug/ml
Target Compounds			
2) N-DECANE	4.502	18804069	50.597 ug/ml
3) N-DODECANE	6.518	18763116	50.341 ug/ml
4) N-TETRADECANE	8.314	18530297	50.037 ug/ml
5) N-HEXADECANE	9.915	18430136	49.733 ug/ml
6) N-OCTADECANE	11.358	18506927	49.375 ug/ml
7) N-EICOSANE	12.666	18153154	48.870 ug/ml
8) N-DOCOSANE	13.863	17815536	48.166 ug/ml
10) N-TETRACOSANE	14.966	17567802	47.329 ug/ml
11) N-HEXACOSANE	15.986	16984009	46.383 ug/ml
12) N-OCTACOSANE	16.935	16640458	45.530 ug/ml
13) N-TRIACONTANE	17.821	16342423	44.895 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062420\
Data File : FE032247.D
Signal(s) : FID1B.CH
Acq On : 24 Jun 2020 19:22
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jun 25 01:14:38 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE061120.M
Quant Title :
QLast Update : Thu Jun 11 13:38:20 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

