

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062422\
 Data File : FE036174.D
 Signal(s) : FID1B.ch
 Acq On : 24 Jun 2022 08:54
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jun 24 23:13:19 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE062022.M
 Quant Title :
 QLast Update : Tue Jun 21 02:15:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.960	6866192	48.931 ug/ml
Target Compounds			
1) N-OCTANE	1.983	7544717	47.163 ug/ml
2) N-DECANE	4.515	8029438	49.340 ug/ml
3) N-DODECANE	6.684	7991303	49.067 ug/ml
4) N-TETRADECANE	8.509	7969388	49.105 ug/ml
5) N-HEXADECANE	10.115	7976924	49.192 ug/ml
6) N-OCTADECANE	11.556	8033959	49.103 ug/ml
7) N-EICOSANE	12.865	7883868	49.134 ug/ml
8) N-DOCOSANE	14.063	7807090	49.093 ug/ml
10) N-TETRACOSANE	15.165	7704163	49.288 ug/ml
11) N-HEXACOSANE	16.184	7541815	49.109 ug/ml
12) N-OCTACOSANE	17.134	7457807	49.571 ug/ml
13) N-TRIACONTANE	18.021	7466469	50.298 ug/ml
14) N-DOTRIACONTANE	18.853	7099453	50.321 ug/ml
15) N-TETRATRIACONTANE	19.637	6175089	48.005 ug/ml
16) N-HEXATRIACONTANE	20.375	4829518	43.378 ug/ml
17) N-OCTATRIACONTANE	21.104	3818322	38.932 ug/ml
18) N-TETRACONTANE	21.996	3597045	36.737 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062422\
Data File : FE036174.D
Signal(s) : FID1B.ch
Acq On : 24 Jun 2022 08:54
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jun 24 23:13:19 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE062022.M
Quant Title :
QLast Update : Tue Jun 21 02:15:59 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

