

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062422AL\
 Data File : FE036216.D
 Signal(s) : FID1B.ch
 Acq On : 25 Jun 2022 08:30
 Operator : YP\AJ
 Sample : N3478-02
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 RO-364

Integration File: autoint1.e
 Quant Time: Jun 25 11:15:01 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 062022.M
 Quant Title : GC Extractables
 QLast Update : Tue Jun 21 03:08:45 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.875	9125568	45.675 ug/ml
Spiked Amount 50.000		Recovery =	91.35%
12) S 1-chlorooctadecane (S...	13.313	7245643	46.801 ug/ml
Spiked Amount 50.000		Recovery =	93.60%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062422AL\
Data File : FE036216.D
Signal(s) : FID1B.ch
Acq On : 25 Jun 2022 08:30
Operator : YP\AJ
Sample : N3478-02
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
RO-364

Integration File: autoint1.e
Quant Time: Jun 25 11:15:01 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 062022.M
Quant Title : GC Extractables
QLast Update : Tue Jun 21 03:08:45 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.875 min
Delta R.T.: 0.004 min
Response: 9125568
Conc: 45.67 ug/ml

Instrument :
FID_E
ClientSampleId :
RO-364



#12 1-chlorooctadecane (SURR)

R.T.: 13.313 min
Delta R.T.: 0.004 min
Response: 7245643
Conc: 46.80 ug/ml