

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062920\
 Data File : FE032272.D
 Signal(s) : FID1B.CH
 Acq On : 29 Jun 2020 12:12
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jun 29 15:15:37 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE061120.M
 Quant Title :
 QLast Update : Thu Jun 11 13:38:20 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.771	15190199	47.510 ug/ml
Target Compounds			
2) N-DECANE	4.506	18750962	50.454 ug/ml
3) N-DODECANE	6.523	18671450	50.095 ug/ml
4) N-TETRADECANE	8.320	18429966	49.767 ug/ml
5) N-HEXADECANE	9.922	18323338	49.445 ug/ml
6) N-OCTADECANE	11.366	18410993	49.119 ug/ml
7) N-EICOSANE	12.676	18080703	48.675 ug/ml
8) N-DOCOSANE	13.874	17768004	48.038 ug/ml
10) N-TETRACOSANE	14.978	17573186	47.343 ug/ml
11) N-HEXACOSANE	15.998	17001202	46.430 ug/ml
12) N-OCTACOSANE	16.947	16576805	45.355 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062920\
Data File : FE032272.D
Signal(s) : FID1B.CH
Acq On : 29 Jun 2020 12:12
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jun 29 15:15:37 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE061120.M
Quant Title :
QLast Update : Thu Jun 11 13:38:20 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

