

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_E\Data\FE070623AL\  
Data File : FE043111.D  
Signal(s) : FID1B.ch  
Acq On : 06 Jul 2023 17:01  
Operator : YP\AJ  
Sample : 03514-04D  
Misc :  
ALS Vial : 17 Sample Multiplier: 1

Instrument :  
FID\_E  
ClientSampleId :  
CP-5

Integration File: autoint1.e  
Quant Time: Jul 06 17:28:38 2023  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_E\methods\Aliphatic EPH 062123.M  
Quant Title : GC Extractables  
QLast Update : Wed Jun 21 11:03:59 2023  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um

| Compound                       | R.T.   | Response   | Conc Units   |
|--------------------------------|--------|------------|--------------|
| -----                          |        |            |              |
| System Monitoring Compounds    |        |            |              |
| 9) S ortho-Terphenyl (SURR)    | 11.905 | 6615573    | 36.012 ug/ml |
| Spiked Amount 50.000           |        | Recovery = | 72.02%       |
| 12) S 1-chlorooctadecane (S... | 13.341 | 5366009    | 39.429 ug/ml |
| Spiked Amount 50.000           |        | Recovery = | 78.86%       |

Target Compounds

-----

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_E\Data\FE070623AL\  
Data File : FE043111.D  
Signal(s) : FID1B.ch  
Acq On : 06 Jul 2023 17:01  
Operator : YP\AJ  
Sample : 03514-04D  
Misc :  
ALS Vial : 17 Sample Multiplier: 1

Instrument :  
FID\_E  
ClientSampleId :  
CP-5

Integration File: autoint1.e  
Quant Time: Jul 06 17:28:38 2023  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_E\methods\Aliphatic EPH 062123.M  
Quant Title : GC Extractables  
QLast Update : Wed Jun 21 11:03:59 2023  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.905 min  
Delta R.T.: -0.021 min  
Response: 6615573  
Conc: 36.01 ug/ml

Instrument :  
FID\_E  
ClientSampleId :  
CP-5



#12 1-chlorooctadecane (SURR)

R.T.: 13.341 min  
Delta R.T.: -0.022 min  
Response: 5366009  
Conc: 39.43 ug/ml