

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE071122\
 Data File : FE036448.D
 Signal(s) : FID1B.ch
 Acq On : 11 Jul 2022 15:28
 Operator : YP\AJ
 Sample : N3214-06
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 12 01:05:40 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE062022.M
 Quant Title :
 QLast Update : Tue Jun 21 02:15:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.946	2690588	19.174 ug/ml
Target Compounds			
1) N-OCTANE	1.967	153110	0.957 ug/ml
2) N-DECANE	4.498	141016	0.867 ug/ml
3) N-DODECANE	6.668	126762	0.778 ug/ml
4) N-TETRADECANE	8.494	120513	0.743 ug/ml
5) N-HEXADECANE	10.100	121044	0.746 ug/ml
6) N-OCTADECANE	11.541	123501	0.755 ug/ml
7) N-EICOSANE	12.850	122787	0.765 ug/ml
8) N-DOCOSANE	14.047	129589	0.815 ug/ml
10) N-TETRACOSANE	15.148	144458	0.924 ug/mlm
11) N-HEXACOSANE	16.167	131065	0.853 ug/ml
12) N-OCTACOSANE	17.116	131971	0.877 ug/ml
13) N-TRIACONTANE	18.003	153795	1.036 ug/ml
14) N-DOTRIACONTANE	18.834	162181	1.150 ug/ml
15) N-TETRATRIACONTANE	19.618	130366	1.013 ug/mlm
16) N-HEXATRIACONTANE	20.357	96547	0.867 ug/ml
17) N-OCTATRIACONTANE	21.080	102821	1.048 ug/mlm
18) N-TETRACONTANE	21.959	102232	1.044 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE071122\
Data File : FE036448.D
Signal(s) : FID1B.ch
Acq On : 11 Jul 2022 15:28
Operator : YP\AJ
Sample : N3214-06
Misc :
ALS Vial : 9 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 12 01:05:40 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE062022.M
Quant Title :
QLast Update : Tue Jun 21 02:15:59 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

