

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE071222\
 Data File : FE036462.D
 Signal(s) : FID1B.ch
 Acq On : 12 Jul 2022 13:48
 Operator : YP\AJ
 Sample : N3214-07
 Misc : 2.5 PPM WATER LOD
 ALS Vial : 13 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 13 01:21:38 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE062022.M
 Quant Title :
 QLast Update : Tue Jun 21 02:15:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.943	2722521	19.402 ug/ml
Target Compounds			
2) N-DECANE	4.496	413787	2.543 ug/ml
3) N-DODECANE	6.666	401484	2.465 ug/ml
4) N-TETRADECANE	8.491	410746	2.531 ug/ml
5) N-HEXADECANE	10.097	405786	2.502 ug/ml
6) N-OCTADECANE	11.538	409282	2.502 ug/ml
7) N-EICOSANE	12.846	405698	2.528 ug/ml
8) N-DOCOSANE	14.044	404748	2.545 ug/ml
10) N-TETRACOSANE	15.146	393273	2.516 ug/ml
11) N-HEXACOSANE	16.165	396955	2.585 ug/ml
12) N-OCTACOSANE	17.114	400687	2.663 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE071222\
Data File : FE036462.D
Signal(s) : FID1B.ch
Acq On : 12 Jul 2022 13:48
Operator : YP\AJ
Sample : N3214-07
Misc : 2.5 PPM WATER LOD
ALS Vial : 13 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 13 01:21:38 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE062022.M
Quant Title :
QLast Update : Tue Jun 21 02:15:59 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

