

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE071520\
 Data File : FE032368.D
 Signal(s) : FID1B.CH
 Acq On : 15 Jul 2020 14:33
 Operator : DD\AJ
 Sample : PB130219BS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 16 01:09:33 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE070720.M
 Quant Title :
 QLast Update : Wed Jul 08 12:04:06 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	14.752	8569090	28.167 ug/ml
Target Compounds				
2)	N-DECANE	4.493	8944741	25.720 ug/ml
3)	N-DODECANE	6.507	9041367	25.908 ug/ml
4)	N-TETRADECANE	8.302	8979493	25.839 ug/ml
5)	N-HEXADECANE	9.903	8950419	25.672 ug/ml
6)	N-OCTADECANE	11.346	8978358	25.390 ug/ml
7)	N-EICOSANE	12.656	8778436	24.960 ug/ml
8)	N-DOCOSANE	13.854	8577475	24.410 ug/ml
10)	N-TETRACOSANE	14.956	8420331	23.788 ug/ml
11)	N-HEXACOSANE	15.976	8113025	23.120 ug/ml
12)	N-OCTACOSANE	16.926	7928087	22.542 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE071520\
Data File : FE032368.D
Signal(s) : FID1B.CH
Acq On : 15 Jul 2020 14:33
Operator : DD\AJ
Sample : PB130219BS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 16 01:09:33 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE070720.M
Quant Title :
QLast Update : Wed Jul 08 12:04:06 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

