

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE071621\
 Data File : FE034071.D
 Signal(s) : FID1B.CH
 Acq On : 16 Jul 2021 10:45
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Jul 17 00:52:01 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE062321.M
 Quant Title :
 QLast Update : Thu Jun 24 03:12:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.693	17674818	50.524 ug/ml
Target Compounds			
1) N-OCTANE	2.386	18020445	48.489 ug/ml
2) N-DECANE	4.436	19256420	50.874 ug/ml
3) N-DODECANE	6.447	19734142	51.554 ug/ml
4) N-TETRADECANE	8.243	20133788	51.437 ug/ml
5) N-HEXADECANE	9.844	20124379	51.411 ug/ml
6) N-OCTADECANE	11.286	20685069	51.320 ug/ml
7) N-EICOSANE	12.597	20637375	51.148 ug/ml
8) N-DOCOSANE	13.796	20443818	50.706 ug/ml
10) N-TETRACOSANE	14.900	20444507	50.239 ug/ml
11) N-HEXACOSANE	15.921	20100262	49.552 ug/ml
12) N-OCTACOSANE	16.872	19705812	48.643 ug/ml
13) N-TRIACONTANE	17.761	19259032	47.624 ug/ml
14) N-DOTRIACONTANE	18.594	18229827	45.996 ug/ml
15) N-TETRATRIACONTANE	19.378	17061651	44.856 ug/ml
16) N-HEXATRIACONTANE	20.117	15243109	43.585 ug/ml
17) N-OCTATRIACONTANE	20.855	13389689	42.922 ug/ml
18) N-TETRACONTANE	21.770	12976885	41.704 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE071621\
Data File : FE034071.D
Signal(s) : FID1B.CH
Acq On : 16 Jul 2021 10:45
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Jul 17 00:52:01 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE062321.M
Quant Title :
QLast Update : Thu Jun 24 03:12:19 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

